

Delft University of Technology

Parameter retrieval of small particles in dark-field Fourier ptychography and a rectangle in real-space ptychography

Wei, Xukang; Urbach, H. Paul; van der Walle, Peter; Coene, Wim M.J.

DOI 10.1016/j.ultramic.2021.113335

Publication date 2021 **Document Version** Final published version

Published in Ultramicroscopy

Citation (APA) Wei, X., Urbach, H. P., van der Walle, P., & Coene, W. M. J. (2021). Parameter retrieval of small particles in dark-field Fourier ptychography and a rectangle in real-space ptychography. *Ultramicroscopy, 229*, Article 113335. https://doi.org/10.1016/j.ultramic.2021.113335

Important note

To cite this publication, please use the final published version (if applicable). Please check the document version above.

Copyright Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

Takedown policy

Please contact us and provide details if you believe this document breaches copyrights. We will remove access to the work immediately and investigate your claim.

Contents lists available at ScienceDirect

Ultramicroscopy



journal homepage: www.elsevier.com/locate/ultramic

Parameter retrieval of small particles in dark-field Fourier ptychography and a rectangle in real-space ptychography



Xukang Wei^{a,*}, H. Paul Urbach^a, Peter van der Walle^b, Wim M.J. Coene^{a,c}

^a Optics Research Group, Delft University of Technology, Lorentzweg 1, Delft, 2628 CJ, The Netherlands ^b TNO Optics Department, Stieltjesweg 1, Delft, 2628 CK, The Netherlands

^c ASML Netherlands B.V, De Run 6501, Veldhoven, 5504 DR, The Netherlands

ARTICLE INFO

Keywords: Ptychograhy Parameterization Cramér Rao Lower Bound Poisson noise

ABSTRACT

We present a parameter retrieval method which incorporates prior knowledge about the object into ptychography. The proposed method is applied to two applications: (1) parameter retrieval of small particles from Fourier ptychographic dark field measurements; (2) parameter retrieval of a rectangular structure with realspace ptychography. The influence of Poisson noise is discussed in the second part of the paper. The Cramér Rao Lower Bound in both applications is computed and Monte Carlo analysis is used to verify the calculated lower bound. With the computation results we report the lower bound for various noise levels and analyze the correlation of particles in application 1. For application 2 the correlation of parameters of the rectangular structure is discussed.

1. Introduction

Ptychography [1–6] is a scanning coherent diffraction imaging method for reconstructing a complex valued object function from intensity measurements recorded in the Fraunhofer or Fresnel diffraction region. In ptychography the object is partially illuminated multiple times with varying position of the illumination spot, so that the entire object is covered and adjacent illuminations partially overlap [7]. The technique provides a solution to the so-called 'phase problem' and is found to be very suitable for EUV [8,9] and X-ray imaging applications [10–13] due to its high fidelity and its minimum requirement on optical imaging elements. Moreover, abundant studies show that ptychography is able to provide a wide field-of-view and retrieve the illumination probe also [14,15]. During the last two decades, ptychography has been successfully demonstrated with X-ray radiation sources [11,16,17], electron beams [18] and visible light sources [19].

More recently, Fourier ptychographic microscopy [20,21] has been proposed, which can be regarded as an extension of ptychography [22]. The technique overcomes the resolution limit of conventional microscopy by enlarging the effective cut-off spatial frequency in the pupil plane. This is done by applying several plane wave illuminations to the sample. The detector is in the image conjugate to the sample plane, and each measurement corresponds to a particular incident angle of the illumination. With each tilted illumination, the diffraction pattern of the sample is shifted in the plane of the exit pupil of the lens, over the aperture used for imaging. Consecutive illumination tilts generate partially overlapping diffraction patterns within the aperture. With all of the Fourier ptychographic measurements, the spatial spectrum of the sample can be synthesized by using ptychographic algorithms with interchanged real space and reciprocal space coordinates [23–25].

In general, the framework of real-space ptychograghy can be described as follows. Let \mathbf{r} and \mathbf{k} be 3D coordinates in real space and reciprocal space:

$$\mathbf{r} = [x, y, z]^T = \begin{bmatrix} \mathbf{r}_{\perp}, z \end{bmatrix}^T, \quad \mathbf{k} = \begin{bmatrix} k_x, k_y, k_z \end{bmatrix}^T = \begin{bmatrix} \mathbf{k}_{\perp}, k_z \end{bmatrix}^T.$$
(1)

and $O(\mathbf{r}_{\perp})$ the object transmission function. We use a laterally shifted probe, denoted by $P(\mathbf{r}_{\perp})$, to illuminate the object multiple times. For the *j*th illumination, the exit wave immediately behind the object is:

$$\Psi_{j}(\mathbf{r}_{\perp}) = P(\mathbf{r}_{\perp} - \mathbf{R}_{\perp,j}) \cdot O(\mathbf{r}_{\perp}) = P_{j}(\mathbf{r}_{\perp}) \cdot O(\mathbf{r}_{\perp}), \qquad (2)$$

where $\mathbf{R}_{\perp,j}$ specifies the shift of the *j*th illumination. The probe function is assumed to have a finite support with, for instance, a circular boundary:

$$P(\mathbf{r}_{\perp}) = \begin{cases} P(\mathbf{r}_{\perp}), & |\mathbf{r}_{\perp}| \le r_0, \\ 0, & |\mathbf{r}_{\perp}| > r_0. \end{cases}$$
(3)

For a detector located at distance *z* in the far field, the diffraction intensity pattern $I(\mathbf{r}')$ for the *j*th illumination is:

$$I_{j}(\mathbf{r}_{\perp}') = \left| \iint \Psi_{j}(\mathbf{r}_{\perp}) \exp\left(-\mathrm{i}\frac{2\pi}{\lambda z}\mathbf{r}_{\perp} \cdot \mathbf{r}_{\perp}'\right) d\mathbf{r}_{\perp} \right|^{2}$$

* Corresponding author. *E-mail address:* x.wei-2@tudelft.nl (X. Wei).

https://doi.org/10.1016/j.ultramic.2021.113335

Received 17 December 2020; Received in revised form 8 March 2021; Accepted 18 May 2021 Available online 24 June 2021

0304-3991/© 2021 The Author(s). Published by Elsevier B.V. This is an open access article under the CC BY license (http://creativecommons.org/licenses/by/4.0/).

$$= \left| \mathcal{F} \left(\Psi_j \right) \left(\mathbf{k}_{\perp}' \right) \right|^2. \tag{4}$$

where \mathcal{F} is the Fourier transform operator. \mathbf{r}'_{\perp} is 2D coordinate in the detector plane. The relation between \mathbf{r}'_{\perp} and \mathbf{k}'_{\perp} is: $\mathbf{k}'_{\perp} = 2\pi \mathbf{r}'_{\perp} (\lambda z)^{-1}$.

The task of ptychography is to find an estimate of the object which fits the given *a priori* knowledge, while a cost function \mathcal{E} is minimized. For the case of real-space ptychography, the *a priori* knowledge is the exact information of the probe function and the set of relative positions \mathbf{R}_j . The cost function \mathcal{E} is defined as the l_2 -distance between the modulus of the far field diffraction pattern $\left|\mathcal{F}\left(\Psi_j\right)(\mathbf{k}_{\perp}')\right|$ and the square root of the measured intensity $I_i^{\mathrm{m}}(\mathbf{k}_{\perp}')$:

$$\mathcal{E} = \sum_{j} \mathcal{E}_{j} = \sum_{j} \sum_{\mathbf{k}_{\perp}'}^{N_{x}^{\text{det}}, N_{y}^{\text{det}}} \left[\sqrt{I_{j}^{\mathbf{m}}(\mathbf{k}_{\perp}')} - \left| \mathcal{F} \left(\Psi_{j} \right) (\mathbf{k}_{\perp}') \right| \right]^{2},$$
(5)

where N_x^{det} and N_y^{det} are the number of pixels of the detector in *x*-axis and *y*-axis, respectively. One way to minimize \mathcal{E} is to use the gradient descent method. If we apply the gradient descent method to each \mathcal{E}_j sequentially, the algorithm is equivalent to the ptychography iterative engine (PIE) [5,6]. Another popular choice is the difference map algorithm, which is formulated in terms of finding the intersection of two constraint sets [14,26]. When the ptychographic measurements contain a relatively large amount of noise, one can utilize de-noising ptychographic algorithms to obtain a better image of the object. One of the most powerful and robust de-noising methods is the Maximum Likelihood estimation [23,27–29], which requires knowledge of the noise model.

The ptychographic measurement $I_j(\mathbf{k}'_{\perp})$ is commonly recorded by a 2D detector, e.g. a charge-coupled device (CCD). Therefore \mathbf{k}'_{\perp} is a discretized grid and is meshed according to the distance *z* and the size of pixel of the detector. The retrieved object function, denoted by \hat{O} , is also on a discretized grid \mathbf{r}_{\perp} . \mathbf{r}_{\perp} and \mathbf{k}'_{\perp} have the relation:

$$[\Delta x, \Delta y]^{T} = 2\pi \left[(N_{x}^{\det} \Delta k_{x}')^{-1}, (N_{y}^{\det} \Delta k_{y}')^{-1} \right]^{T},$$
(6)

where Δx and Δy are the spacing of a single grid cell in *x*-axis and *y*-axis, respectively, and $\Delta k'_x$ and $\Delta k'_y$ are the spacing of a grid cell in k_x and k_y , respectively. Note that the total field-of-view (FoV) in the object plane is:

$$FoV = \left[N_x \Delta x, N_x \Delta y\right]^T, \tag{7}$$

where $N_x > N_x^{\text{det}}$ and $N_y > N_y^{\text{det}}$ due to that ptychography is a scanning imaging technique which provides an extended FoV. In line with this extended FoV, we have the effective spacing of the grid cell in the reciprocal space:

$$\left[\Delta k_x, \Delta k_y\right]^T = \left[\left(N_x\right)^{-1} N_x^{\det} \Delta k'_x, \left(N_y\right)^{-1} N_y^{\det} \Delta k'_y\right]^T.$$
(8)

We can see that, when the influence of noise is negligible, the relation given in Eq. (6) imposes a resolution limit to the reconstruction in ptychography. To overcome this limit, several 'superresolution' methods have been proposed [30–32]. One of the ideas lying behind these methods is to impose additional *a priori* knowledge about the object, e.g. analytic continuation of the Fourier transform of bounded support [33–35] or sparsity [31,32], to the algorithm. In this paper we show a parameter retrieval algorithm which incorporates additional *a priori* knowledge about the object into ptychography. We present this algorithm by numerically demonstrating two applications:

(1) Parameter retrieval of sub-wavelength particles using Fourier ptychography with dark field measurements only. For this example the configuration is in line with the 'RapidNano' particle scanner developed by TNO [36,37]. The particle scanner is supposed to detect nano-particles on an EUV reticle. Since only dark field images are recorded in the scanner, the part of the spatial spectrum of the object in the neighborhood of $|\mathbf{k}_{\perp}| = 0$ is lost. The missing data can in principle be filled in by analytic

continuation using the fact that the object has bounded support, however, this method is unstable with noisy measurement and leads in practice to incorrect reconstructions [35,38]. However, as shown in Section 2, the proposed parameter retrieval algorithm is able to extract information of sub-wavelength particles from the incomplete data.

(2) Parameter retrieval of rectangular objects using real-space ptychography. This example comes from practical applications in semiconductor industry where we often want to measure the transmission, the width and the position of the rectangles on flat substrates [39,40]. We demonstrate the proposed parameter retrieval method for this application in Section 3.

To study the influence of Poisson noise on the proposed parameter retrieval scheme, we compute the Cramér Rao Lower Bound (CRLB) and perform Monte Carlo analysis for both two applications in the second part of this paper. We derive the general form of the Fisher information matrix in Section 4. For application 1, the calculated CRLB and Monte Carlo result are shown in Section 5. For application 2, the discussion about the correlation of the parameters of the rectangular structure can be found in Section 6. For both applications, it was found that the uncertainty of the parameter retrieval is inversely proportional to the photon counts, and potentially is not limited by the sizes of individual cells of the discretized meshgrid in object space.

2. Application 1: parameter retrieval of sub-wavelength particles using fourier ptychography with dark field measurement

2.1. Description of the 'RapidNano 3' particle scanner

The 'RapidNano 3' particle scanner [36,37] is designed to detect small dielectric particles on a flat substrate. The particles are made of polystyrene latex (PSL) beads and the typical diameter of the particle is ~ 50 nm. The scanner has detection limit of 42 nm PSL particles, i.e. the capture rate is 95% at this size. Note that the particles on the substrate can be any material and PSL is only the calibration standard. The particles are sparsely distributed on the sample mostly. The substrate is reflective, made of silicon, and its lateral size can be up to 6×6 inch, i.e. the size of an EUV mask. The illumination is a 532 nm, *p*-polarized, fully coherent plane wave laser beam. The incident angle of the illumination is 60 degree, with 9 regularly distributed azimuth incident directions around 360 degree. The NA of the objective lens is 0.4, therefore the measurement is a dark field image of the sample as is illustrated in Fig. 1.

2.2. Single dipole radiation

Considering that the diameter of the detected particles is around 10 times smaller than the illumination wavelength, we begin by using the dipole radiation formula to model the wavefield scattered by the particles. Suppose that there are *N* dipoles in the plane z = 0, and the *i*th oscillating dipole is located at position $\mathbf{r}_i = [\mathbf{r}_{\perp,i}, 0]^T$, i = 1, 2, ..., N, and is excited by an incident plane wave $\mathbf{E}_{\text{in},i}$:

$$\mathbf{E}_{\text{in},j} = A_{\text{in}} e^{\mathbf{i}\mathbf{k}_j \cdot \mathbf{r}} \hat{\mathbf{e}}_p(\mathbf{k}_j) = A_{\text{in}} e^{\mathbf{i}\mathbf{k}_{\perp,j} \cdot \mathbf{r}_{\perp}} \hat{\mathbf{e}}_p(\mathbf{k}_j), \tag{9}$$

where A_{in}^2 is proportional to the illumination power and $\hat{\mathbf{e}}_p(\mathbf{k}_j)$ denotes the polarization direction.

For the *i*th dipole with position $\mathbf{r}_{\perp,i}$, we denote the dipole moment by:

$$\mathbf{p}_{i,j} = \alpha_i \mathbf{E}_{\mathrm{in},j} = \alpha_i A_{\mathrm{in}} e^{\mathbf{i}\mathbf{k}_{\perp,j}\cdot\mathbf{r}_{\perp,i}} \hat{\mathbf{e}}_p(\mathbf{k}_j), \tag{10}$$

where ϵ_0 is the permittivity of free space and α_i is the polarizability of the particle. For a dielectric sphere with diameter *d*, the dipole moment $\mathbf{p}_{i,i}^{\text{sphere}}$ in the quasi-static approximation is given by:

$$\mathbf{p}_{i,j}^{\text{sphere}} = \left(\frac{\epsilon_r - 2}{\epsilon_r + 1}\right) d_i^3 \mathbf{E}_{\text{in},j},\tag{11}$$

where $\epsilon_r = n_{\rm PSL}^2$ is the relative permittivity of the dielectric. $n_{\rm PSL}$ is the refractive index of the small particles. Since the real part of $n_{\rm PSL}$ is $\sim 10^6$ times larger than the imaginary part, i.e. than the absorption index, we assume the α_i is real valued for the rest of this paper. We see that α_i is proportional to the volume of the dielectric particle.

The electric field radiating from the *i*th dipole due to the *j*th illumination is given by [41,42]:

$$\mathbf{E}_{\text{scat},i,j} = \mathbf{\vec{G}} \left(\mathbf{r}, \mathbf{r}_i \right) \mathbf{p}_{i,j}, \tag{12}$$

where $\mathbf{G}(\mathbf{r}, \mathbf{r}_i)$ is the dyadic Green's function:

$$\stackrel{\leftrightarrow}{\mathbf{G}}(\mathbf{r},\mathbf{r}_{i}) = \frac{k^{2}}{4\pi\epsilon_{0}} \left(\stackrel{\leftrightarrow}{\mathbf{I}} + \frac{1}{k^{2}}\nabla\nabla\right) \frac{e^{ik|\mathbf{r}-\mathbf{r}_{i}|}}{|\mathbf{r}-\mathbf{r}_{i}|},$$
(13)

where \mathbf{I} is the 3 × 3 identity matrix. Considering that the detector of the particle scanner is insensitive to the polarization state and that the NA of the objective lens is 0.4, we ignore the effect of the polarization of the wavefield for simplicity. Hence we arrive at a scalar scattered amplitude given by:

$$E_{\text{scat},i,j} = A_{\text{in}} k^2 \alpha_i e^{i\mathbf{k}_{\perp,j} \cdot \mathbf{r}_{\perp,i}} G\left(\mathbf{r}, \mathbf{r}_i\right), \qquad (14)$$

where

$$G(\mathbf{r},\mathbf{r}_{i}) = \frac{k^{2}}{4\pi\epsilon_{0}} \frac{e^{ik|\mathbf{r}-\mathbf{r}_{i}|}}{|\mathbf{r}-\mathbf{r}_{i}|}.$$
(15)

2.3. Dark field measurement from the particle scanner

By Fourier transforming Eq. (14) with respect to \mathbf{r}_{\perp} , we have:

$$\mathcal{F}\left(E_{\text{scat},i,j}\right)(\mathbf{k}_{\perp},z) = A_{\text{in}}k^{2}\frac{e^{ik_{z}|z|}}{8i\pi\epsilon_{0}k_{z}}\alpha_{i}e^{-i\mathbf{r}_{\perp,i}\cdot(\mathbf{k}_{\perp}-\mathbf{k}_{\perp,j})}.$$
(16)

 $\mathcal{F}(E_{\text{scat},i,j})$ can be regarded as the 2D spatial spectrum of the scattered wavefield in the plane *z*. The wavefield passes through the imaging system and forms an image in the image plane. The imaging system acts as a low-pass filter as given in [43]. The detector measures the low-pass filtered image of the sample, i.e. the field in the plane $z \to 0$. We therefore let $e^{ik_z|z|} \approx 1$ for this moment. The low-passed wavefield at the exit pupil is given by:

$$\mathcal{F}\left(E_{\text{scat},i,j}\right)_{\text{exit}}(\mathbf{k}_{\perp}) = \mathbf{1}_{k\text{NA}}(\mathbf{k}_{\perp})\frac{A_{\text{in}}k^2}{8i\pi\epsilon_0 k_z}\alpha_i e^{-i\mathbf{r}_{\perp,i}\cdot\left(\mathbf{k}_{\perp}-\mathbf{k}_{\perp,j}\right)},\tag{17}$$

where $\mathbf{1}_{kNA}(\mathbf{k}_{\perp})$ represents the numerical aperture of the objective lens:

$$\mathbf{1}_{k\mathrm{NA}}(\mathbf{k}_{\perp}) = \begin{cases} 1, & |\mathbf{k}_{\perp}| \le k\mathrm{NA}, \\ 0, & |\mathbf{k}_{\perp}| > k\mathrm{NA}. \end{cases}$$
(18)

By summing over all the dipoles, we find the total field in the exit pupil, which is denoted by Ψ_i :

$$\Psi_{j}(\mathbf{k}_{\perp}) = \sum_{i} \mathcal{F} \left(E_{\text{scat},i,j} \right)_{\text{exit}} (\mathbf{k}_{\perp})$$

$$= \mathbf{1}_{k\text{NA}}(\mathbf{k}_{\perp}) \frac{A_{\text{in}}k^{2}}{8i\pi\epsilon_{0}k_{z}} \sum_{i} \alpha_{i} e^{-i\mathbf{r}_{\perp,i} \cdot (\mathbf{k}_{\perp} - \mathbf{k}_{\perp,j})}$$

$$= Q(\mathbf{k}_{\perp}) \cdot O(\mathbf{k}_{\perp} - \mathbf{k}_{\perp,j}), \qquad (19)$$

where

$$Q(\mathbf{k}_{\perp}) = \mathbf{1}_{k\mathrm{NA}}(\mathbf{k}_{\perp}) \frac{A_{\mathrm{in}}k^2}{8i\pi\epsilon_0 k_z},$$
(20)

and $\mathit{O}(\mathbf{k}_{\perp})$ is the Fourier transform of the object defined by

$$O(\mathbf{k}_{\perp}) = \sum_{i} \alpha_{i} e^{-i\mathbf{k}_{\perp} \cdot \mathbf{r}_{\perp,i}}.$$
(21)

Note that the object function is assumed to be independent of the angle of incidence, i.e. the only effect of the tilted illumination is the shift of the Fourier transform of the object function over the pupil plane. Finally, by inverse Fourier transforming Ψ_i and taking the squared



Fig. 1. Graphical illustration of $O(\mathbf{k}_{\perp})$. (a) The blue disk is defined by $\mathbf{1}_{k\rm NA}(\mathbf{k}_{\perp})$ and indicates information about *O* included in the single measurement $I_j(\mathbf{r}'_{\perp})$. (b) The retrievable part of *O* from all given dark field measurements. (For interpretation of the references to color in this figure legend, the reader is referred to the web version of this article.)

modulus, we arrive at the expression for the measured intensity in the detector plane:

$$I_{j}(\mathbf{r}_{\perp}') = \left| \mathcal{F}^{-1} \left(\Psi_{j} \right) \right|^{2} (\mathbf{r}_{\perp}')$$
$$= \left| \mathcal{F}^{-1} \left[Q(\mathbf{k}_{\perp} + \mathbf{k}_{\perp,j}, z) \cdot O(\mathbf{k}_{\perp}) \right] \right|^{2} (\mathbf{r}_{\perp}'), \tag{22}$$

where \mathbf{r}'_{\perp} is the 2D regular grid.

For the current configuration of the particle scanner, $|\mathbf{k}_{\perp,j}|$ is fixed and equal to $k \sin(\frac{\pi}{3})$. This incident angle have been chosen in [36,44], where the simulations show that ratio between the scattering amplitude of the particles and the background substrate is maximized. In this paper we keep the value of $|\mathbf{k}_{\perp,j}|$ be $k \sin(\frac{\pi}{3})$, but the value can be varied which is regarded as a potential subject in the future. The NA of the objective lens is ~ 0.4. Therefore, the intensity measurements do not contain any information about $O(\mathbf{k}_{\perp} = 0)$ and its surrounding region, as shown in Fig. 1. The blue shaded area in Fig. 1(a) illustrates the information about $O(\mathbf{k}_{\perp})$ included in the single measurement $I_j(\mathbf{r}'_{\perp})$, while the blue shaded area in Fig. 1(b) represents the retrievable information from all measurements. We denote this retrievable part of O by $\mathbf{1}_{\Omega}O(\mathbf{k}_{\perp})$:

$$\mathbf{1}_{\Omega}(\mathbf{k}_{\perp}) = \begin{cases} 1, & \mathbf{k}_{\perp} \in \Omega, \\ 0, & \mathbf{k}_{\perp} \notin \Omega, \end{cases}$$
(23)

where Ω is the blue shaded region in Fig. 1(b).

2.4. Retrieving the parameters of the particles

To retrieve α_i and the position $\mathbf{r}_{\perp,i}$ of the dipoles, we first reconstruct the complex valued function $\mathbf{1}_{\Omega}O(\mathbf{k}_{\perp})$ in the pupil plane from the set of intensity measurements $I_j(\mathbf{r}'_{\perp})$. This can be done by applying a ptychographic algorithm. For noisy measurements, one may use the Maximum Likelihood estimator (MLE) if one can find a dominant noise model [27,28]. For the case of Poisson noise, we can apply gradient descent methods [45,46] to minimize the likelihood function \mathcal{L}_P given by Eq. (S7) in the Supplement. We use the MLE ptychographic algorithm which is given in [47] in the simulation and use $\mathbf{1}_{\Omega}\hat{O}(\mathbf{k}_{\perp})$ to denote the reconstruction obtained by the ptychographic method.

Once $\mathbf{1}_{\Omega}\hat{O}(\mathbf{k}_{\perp})$ is obtained, we apply the method of least square to estimate α_i and $\mathbf{r}_{\perp,i}$ of all dipoles. The number of freedom in this problem is $N \times 3$, where N is the number of dipoles within the field-of-view (FoV). Note that if α_i is complex valued, the degrees of freedom should be $N \times 4$. When N is in the order of $10^0 \sim 10^1$, we have much less degrees of freedom than in the traditional Fourier ptychography problem.

Our proposed parameter retrieval algorithm is shown in the following.



Fig. 2. Illustration of the setup of application 1. The incident angle of the illumination is 60 degree, with multiple azimuth incident directions around 360 degree.

Table 1

Configuration settings in the simulation.

Illumination			Imaging system		
Wavelength	Incident angle		NA	Magnification	
500 nm	60 degree		0.4	20	
Detector				Grid spacing in object plane	
Pixel size	Pixel number	FoV		Δx and Δy	
5 µm	200×200	50 µm		133.3 nm	

- Use the MLE ptychographic algorithm to retrieve the complex valued wavefield 1_QO(k_⊥) in the pupil plane.
- (2) From all the dark field intensity measurements, find the lower and upper bound of α_i and **r**_{⊥,i} for i = 1, 2, ..., N. These bounds are denoted by: αⁱ_i, αⁱ_i, **r**ⁱ_⊥, and **r**^u_{⊥i}.
- (3) Solve the following problem:

$$\underset{\alpha_{i},\mathbf{r}_{\perp,i}}{\operatorname{arg min}} \left\| \mathbf{1}_{\Omega} \hat{O}(\mathbf{k}_{\perp}) - \sum_{i} \alpha_{i} e^{-i\mathbf{k}_{\perp} \cdot \mathbf{r}_{\perp,i}} \right\|_{\mathbf{k}_{\perp} \in \Omega}^{2},$$
subject to $\alpha_{i}^{l} \leq \alpha_{i} \leq \alpha_{i}^{u}, \quad i = 1, 2, ..., N,$
 $\mathbf{r}_{\perp,i}^{l} \leq \mathbf{r}_{\perp,i} \leq \mathbf{r}_{\perp,i}^{u}, \quad i = 1, 2, ..., N,$

$$(24)$$

where \leq denotes vector inequality: $\mathbf{r}_{\perp}^{l} \leq \mathbf{r}_{\perp}^{u}$ means $x^{l} \leq x^{u}$ and $y^{l} \leq y^{u}$.

2.5. Simulation

To validate the proposed parameter retrieval algorithm, a preliminary simulation is reported in this section. The configuration is drawn in Fig. 2 and the parameter settings of the setup is described in Table 1. Since the NA of the imaging system is smaller than $|\mathbf{k}_{\perp,j}|$, the measurements at the detector plane are always dark field images. We assume that the detector is insensitive to the polarization state of the wavefield.

The simulated sample consists of two dipoles. The actual scattering strength α_i and the position $\mathbf{r}_{\perp,i}$ of the dipoles are listed in Table 2. Based on these given parameters, we first construct the actual complex valued function $\mathbf{1}_{\Omega}O(\mathbf{k}_{\perp})$ according to Eq. (20). The dark field intensity measurements are noise-free and computed in accordance with Eq. (22). In line with the 1st step of the proposed method given in Section 2.4, the reconstructed object function, denoted by $\mathbf{1}_{\Omega}\hat{O}(\mathbf{k}_{\perp})$, is obtained by applying the Fourier ptychography method. We assume that the function $Q(\mathbf{k}_{\perp} + \mathbf{k}_{\perp,j})$ is known and we ignore the polarization state. In the simulation we notice that only 9 incident plane waves cannot provide sufficient data redundancy. The percentage of sufficient data redundancy have been reported in ptychography literature when

the illumination is bounded by circular support [48]. Suppose the diameter of the circular support is *L*, and the distance between adjacent illumination positions is denoted by $d \in [0, L]$. The overlap ratio is defined by: $1 - \frac{d}{L}$, which is usually be assigned from 60% to 85% to achieve optimal performance [48,49]. In this simulation we use 36 plane waves, instead of 9, with regularly distributed azimuth incident directions around 360 degrees, which means the overlap ratio is 81.13%. The actual function *O* and the reconstructed one are shown in Fig. 3(a) and Fig. 3(b), respectively. Fig. 3(c) illustrates the illuminated area in the reciprocal space, i.e. $\sum_j \mathbf{1}_{kNA}(\mathbf{k}_{\perp} + \mathbf{k}_{\perp,j})$, for 9 and 36 dark field measurements, respectively.

In Fig. 4(a) we show the incoherent sum of all 36 simulated noisefree intensity measurements, i.e. $\sum_{j} I_{j}(\mathbf{r}'_{\perp})$, and in Fig. 4(b) we present the squared amplitude of the scattered field from the sample at plane $z \rightarrow 0$, i.e. $\left|\mathcal{F}^{-1}(\mathbf{1}_{\Omega}\hat{O})\right|^{2}(\mathbf{r}_{\perp})$. For application 1 the spacing of grid \mathbf{r}'_{\perp} and \mathbf{r}'_{\perp} fulfills:

$$[\Delta x, \Delta y]^T = \left[(N_x)^{-1} N_x^{\text{det}} \Delta x', (N_y)^{-1} N_y^{\text{det}} \Delta y' \right]^T,$$
(25)

which can be derived from Eq. (8) by interchanging the real space and reciprocal space coordinates. The inserted graphs in Fig. 4 correspond to dipole *i* = 1. Note that every dark field measurement is a 200 × 200 array with a 250 nm pixel size, which is due to the detector has 200 × 200 pixels with 5 µm pixel size and the magnification of the imaging system is 20, as given in Table 1. The reconstructed scattered field shown in Fig. 4(b) only contains information of $\mathbf{k}_{\perp} \in \Omega$. The side-lobe which appears in the neighborhood of the particles in Fig. 4(b) is due to the fact that the reconstruction is convolved by $\mathcal{F}^{-1}(\mathbf{1}_{\Omega})(\mathbf{r}_{\perp})$. Without knowing the wavefield at $\mathbf{k}_{\perp} = 0$ and its surrounding region or without considering any prior information about the sample, the reconstructed scattering field cannot provide a unique physical solution.

Once $\mathbf{1}_{\Omega}\hat{O}(\mathbf{k}_{\perp})$ is obtained, we retrieve α_i and $\mathbf{r}_{\perp,i}$ by minimizing the least square function given in Eq. (24). This is done by using the 'fmincon' solver in MATLAB. To facilitate the solver to find the global minimum, a proper starting search point and a set of bounds for α_i and $\mathbf{r}_{\perp i}$ are needed. From Fig. 4 we see that one can deduce a guess about the scattering strength and the position of the dipoles from the dark field measurements. Based on the guess we can obtain the starting point and the bounds. The accuracy of the guess of the position is limited by the pixel size of the detector. In the simulation we deduce the initial guess as follows. We first choose in Fig. 4(b) one pixel cell which approximately have minimal and equal distances from the centers of the images of two dipoles. In Fig. 4(b) the indices of this pixel cell in the x and y directions are $[195, 186]^T$. Then we set the top left corner of this pixel cell as origin. The initial guess of position of the dipoles are obtained by roughly measuring the distance between the origin and the center of the image of the dipoles in Fig. 4(b). To determine the bounds of the position, we first choose two 5×5 pixel arrays which center at the brightest pixel cells of the image of two dipoles, respectively. We choose to use the 5×5 arrays because the sum of the absolute square of the value of each corresponding 5×5 pixels is approximately equal to 90% of the total scattering intensity of each dipole. The bounds of the positions are determined by the outer boundary of the two 5×5 pixel arrays in the x and y, respectively. The starting search point of the position of each particle is randomly assigned inside the corresponding outer boundaries, respectively. The initial guess of the scattering strength of each dipole, on the other hand, is determined by summing the absolute square of the value over the corresponding 5×5 pixel arrays of each dipole, respectively. The lower bounds of the scattering strength of the dipoles are set to be 0 and the upper bound are left undetermined, i.e. positive infinity. The starting point of all parameters are shown in Table 2. The retrieved parameters are listed in the most right column of the same table.



Fig. 3. (a) The amplitude and phase of the actual complex function $\mathbf{1}_{\Omega}O(\mathbf{k}_{\perp})$. (b) The amplitude and phase of $\mathbf{1}_{\Omega}O(\mathbf{k}_{\perp})$ which is reconstructed from the Fourier ptychographic algorithm. (c) Illustration of the illuminated area in the reciprocal space, i.e. $\sum_{i} \mathbf{1}_{kNA}(\mathbf{k}_{\perp} + \mathbf{k}_{\perp,i})$, for 9 and 36 dark field measurements, respectively.



Fig. 4. (a) The normalized incoherent sum of all 36 dark field measurements, i.e. $\sum_{j} I_{j}(\mathbf{r}'_{\perp}) / \max\left[\sum_{j} I_{j}(\mathbf{r}'_{\perp})\right]$. (b) The normalized amplitude of scattering wavefield at plane $z \to 0$, i.e. $\left|\mathcal{F}^{-1}(\mathbf{1}_{\Omega}\hat{O})\right|^{2}(\mathbf{r}_{\perp}) / \max\left[\left|\mathcal{F}^{-1}(\mathbf{1}_{\Omega}\hat{O})\right|^{2}(\mathbf{r}_{\perp})\right]$, which is reconstructed with the Fourier ptychography method. The inserted graphs correspond to the dipole i = 1.

Table 2									
Retrieved	parameters	of	two	dipoles	in	the	noise	free	simulation

	Actual value	Initial guess	Retrieved value
$\alpha_1/(\lambda^3)$	1.000×10^{-3}	0.935×10^{-3}	1.000×10^{-3}
<i>x</i> ₁	-8.333 μm	-8.349 μm	-8.333 μm
<i>y</i> ₁	0.000 µm	0.113 µm	0.000 µm
$\alpha_2/(\lambda^3)$	0.512×10^{-3}	0.429×10^{-3}	0.512×10^{-3}
x2	8.356 µm	8.327 μm	8.356 µm
<i>y</i> ₂	0.088 µm	-0.029 μm	0.088 µm

3. Application **2**: parameter retrieval of a rectangular object using real-space ptychography

3.1. Single object embedded in constant surrounding

Now we consider a real-space ptychography setup as shown in Fig. 5. The object can be written in the following form:

$$O(\mathbf{r}_{\perp})(A_1, \phi_1, a_1, b_1, \mathbf{r}_{\perp,1}) = 1 + (A_1 e^{i\phi_1} - 1)\Pi_{a_1, b_1}(\mathbf{r}_{\perp} - \mathbf{r}_{\perp,1})$$

= 1 + C_1 \Pi_{a_1, b_1, \mathbf{r}_{\perp,1}}, (26)

where $C_1 = A_1 e^{i\phi_1} - 1$ is a complex valued coefficient. and $\Pi_{a_1,b_1}(\mathbf{r}_{\perp} - \mathbf{r}_{\perp,1})$ is the 2D rectangular function defined by parameters:

$$\Pi_{a_{1},b_{1}}(\mathbf{r}_{\perp} - \mathbf{r}_{\perp,1}) = \Pi_{a_{1}}(x - x_{1})\Pi_{b_{1}}(y - y_{1}) \\
= \begin{cases} 0, & |x - x_{1}| > \frac{a_{1}}{2} \text{ or } |y - y_{1}| > \frac{b_{1}}{2}, \\ 1, & |x - x_{1}| < \frac{a_{1}}{2} \text{ and } |y - y_{1}| < \frac{b_{1}}{2}. \end{cases}$$
(27)



Fig. 5. The configuration of application 2.

We aim to retrieve the parameters:

$$\Theta = \left[A_1, \phi_1, a_1, b_1, \mathbf{r}_{\perp,1}\right]^T, \text{ where: } A_1 \in (0, 1], \ a_1 > 0, \ b_1 > 0.$$
(28)

The diffracted wavefield in the far field for the *j*th illumination is:

$$\mathcal{F}(\Psi_{j})\left(\mathbf{k}_{\perp}^{\prime}\right) = \mathcal{F}\left(P_{j}\right)\left(\mathbf{k}_{\perp}^{\prime}\right) + \mathcal{F}\left(P_{j}\right)\left(\mathbf{k}_{\perp}^{\prime}\right) \otimes \left[C_{1}a_{1}b_{1}\operatorname{sinc}\left(\frac{a_{1}k_{x}}{2}\right)\operatorname{sinc}\left(\frac{b_{1}k_{y}}{2}\right)e^{i\mathbf{k}_{\perp}^{\prime}\cdot\mathbf{r}_{\perp,1}}\right], \quad (29)$$

where \otimes denotes convolution. Note in Eq. (27) we leave the values of the function at $x = x_1 \pm \frac{a_1}{2}$ and $y = y_1 \pm \frac{b_1}{2}$ be undefined because



Fig. 6. (a) The simulated object in Fourier space. The object has one rectangle which is embedded in a constant surrounding. (b) The normalized amplitude and the phase of the probe, which is known in the simulation. (c) The retrieved object function in Fourier space from ptychographic measurement. (d) The inverse Fourier transform of $\mathcal{F}(\hat{O})(\mathbf{k}_{\perp})$.

these values cannot be retrieved under the projection approximation given by Eq. (2). We can see in Eq. (29) that the diffracted wavefield is not a function w.r.t. the value of $O(\mathbf{r}_{\perp})$ at position $x = x_1 \pm \frac{a_1}{2}$ and $y = y_1 \pm \frac{b_1}{2}$. The validity of the projection approximation have been discussed in [2,11] and we assume in the paper that this approximation is valid.

3.2. Retrieving the parameter of the rectangle

We can see in Eq. (29) that, when we have exact knowledge of the probe, the diffraction pattern is a function of the parameters of the rectangle. This fact offers us the chance to retrieve the parameters given in Eq. (28) from the measurements $I_j(\mathbf{k}'_{\perp})$ for all *j*. In this section we propose and validate a feasible method to retrieve the parameters from a ptychographic measurement.

The first step of the proposed method is to reconstruct the object function in real space, denoted by: $\hat{O}(\mathbf{r}_{\perp})$, from $I_j(\mathbf{k}'_{\perp})$ for all *j*. This can be done by applying the PIE [4,5] algorithm or other ptychography algorithms [14,29,50,51]. The discretization of \mathbf{r}_{\perp} and \mathbf{k}'_{\perp} follows Eq. (6). Note that $\hat{O}(\mathbf{r}_{\perp})$ can be obtained even if the probe function is unknown, which is due to the data redundancy of the ptychographic measurement.

Once the minimum of the likelihood function is found, we can compute the Fourier transform of the reconstructed object, denoted by $\mathcal{F}(\hat{O})(\mathbf{k}_{\perp})$. The spacing of grid \mathbf{r}_{\perp} and \mathbf{k}_{\perp} is given in Eq. (8). The parameter of the rectangle can be retrieved by minimizing a cost function \mathcal{G} defined by:

$$\mathcal{G} = \left\| \mathcal{F}\left(\hat{O} - 1\right) - C_1 a_1 b_1 \operatorname{sinc}\left(\frac{a_1 k_x}{2}\right) \operatorname{sinc}\left(\frac{b_1 k_y}{2}\right) e^{\mathbf{i} \mathbf{k}_{\perp} \cdot \mathbf{r}_{\perp,1}} \right\|^2, \quad (30)$$

where $\|\cdot\|^2$ denotes the l_2 norm. To give an example about the relation between \mathcal{G} and the rectangle parameters, we show in Fig. 7 the value of \mathcal{G} as a function of a_1 and x_1 . The configuration parameter of Fig. 7 will be given later in Section 4. It is seen that \mathcal{G} is convex in the neighborhood of the actual a_1 and x_1 , which offers us the chance to retrieve the parameter by minimizing \mathcal{G} . In order to find the minimum of G, it will be beneficial to start the algorithm from a point close to the actual value. This starting point can be determined from $\hat{O}(\mathbf{r})$. In summary, our proposed method includes the following steps:

- Use the MLE ptychographic algorithm to retrieve the complex valued wavefield Ô(r_⊥).
- (2) Find the lower and upper bound of *Θ* from *Ô*(**r**_⊥). *Θ* is the parameter vector defined by Eq. (28). These bounds are denoted by: *Θ_l* and *Θ_u*.
- (3) Solve the following problem:

$$\arg\min \mathcal{G}, \quad \text{subject to } \Theta_l \le \Theta \le \Theta_u. \tag{31}$$

3.3. Simulation

To validate our proposed method, a preliminary simulation is shown. We consider the setup as shown in Fig. 5. Details of the configuration are shown in Table 3. The Fresnel number of this configuration is 0.0014. According to Eq. (29), we first generate the complex valued wavefield in Fourier space $\mathcal{F}(\Psi_j)(\mathbf{k}_{\perp})$ based on the given probe and object. The Fourier transform of the object function $\mathcal{F}(O)(\mathbf{k}_{\perp})$ is illustrated in Fig. 6(a). The object consists of one rectangle with sizes listed in Table 4. Fig. 6(b) shows the normalized amplitude and the phase of the probe. In this simulation we assume the probe is known and the ptychographic measurement is noise-free. In Fig. 6(c) we illustrate the Fourier transform of the reconstructed object function $\mathcal{F}(\hat{O})(\mathbf{k}_{\perp})$. The inverse Fourier transform of $\mathcal{F}(\hat{O})(\mathbf{k}_{\perp})$ is shown in Fig. 6(d).

After obtaining $\mathcal{F}(\hat{O})(\mathbf{k})$, we can retrieve the parameters of the rectangle by solving the optimization problem in Eq. (31). In Fig. 7 we demonstrate the evaluation of the cost function \mathcal{G} with respect to a_1 and x_1 , which are the width and position of the rectangular in the *x*-direction. The orange arrows in both plots points to the actual values of a_1 and x_1 . We see in Fig. 7 that it is possible to accurately retrieve the values of a_1 and x_1 by minimizing \mathcal{G} . To compute the solution of the problem in Eq. (31), we again implemented the 'fmincon' solver in MATLAB. Furthermore, Fig. 7 shows that the value of \mathcal{G} is approximately a quadratic function w.r.t. a_1 and a linear function w.r.t. x_1 in

Ultramicroscopy 229 (2021) 113335

Table 3 The cha

he	characteristic	parameters	of	the	configuration	in	the	simulation.	

Probe	Grid size 60×60	Grid spacing 30 nm	Wavelength 30 nm	Scanning grid 5×5	Overlap ratio 75%	Radius of circular support 0.45 µm
object	Grid size 90 × 90	Grid spacing 30 nm	Detector	Pixel number 60×60	Pixel size 50 μm	Propagation distance 1.88 cm



Fig. 7. The evaluation of \mathcal{E} with respect to a_1 and x_1 . The value of \mathcal{E} is normalized to its maximum in both plots. The orange arrow points to the actual value of a_1 and x_1 in this simulation.

Table 4

Retrieved parameters of one rectangle.

	a_1/λ	b_1/λ	x_1/λ	y_1/λ	A_1	ϕ_1
Actual value	11.46	25.99	5.71	1.42	0.70	3.14
Initial guess	11.00	28.00	4.00	3.00	0.73	3.17
Retrieved value	11.46	25.99	5.71	1.42	0.70	3.14

the neighborhood of the actual values, which is explained in Section 2 of Supplement. The actual value of the parameters, the starting point and the retrieved results are presented in Table 4. We can see that the proposed method can successfully retrieve the parameters of the rectangle.

4. The CRLB analysis of the parameter retrieval scheme for Poisson noise

In estimation theory, the Cramér Rao Lower Bound (CRLB) gives a lower bound on the variance of any unbiased estimator for a parameter that is to be estimated. The estimators that can reach the lower bound are called the minimum variance unbiased estimators. Minimum variance unbiased estimators are often not available [52,53]. To find the CRLB, one needs to compute the Fisher information matrix which is the expectation value of the second order derivative of the likelihood function. Detailed description about CRLB, Fisher information matrix and Maximum Likelihood Estimation is given in Section 1.A of Supplement.

In this paper we study the CRLB for Poisson distributed photon counting noise, which is the most dominant source of noise which naturally occurs even under the best experimental conditions [27,28]. The expectation of the second order derivative of the Poisson likelihood function can be found in Section 1.B of Supplement.

5. The CRLB analysis of application 1

5.1. The Fisher information matrix for retrieval of the dipole

Now we calculate the Fisher matrix for the *i*th dipole. According to Eq. (24), the parameters we aim to estimate are:

$$\Theta = \left[\theta_1, \theta_2, \dots, \theta_N\right]^T = \left[\alpha_1, x_1, y_1, \alpha_2, x_2, y_2, \dots, \alpha_N, x_N, y_N\right]^T.$$
(32)

We consider that we aim to retrieve the parameters of the *i*th dipole while assuming that the parameters of all other dipoles are known. To find the Fisher matrix, we need to calculate the derivative of I_j with respect to the parameters of dipole *i*. The derivatives of I_j are given

in Section 1.C of the Supplement. The number of elements of I_j are determined by the amount of dipoles. For the case of two dipoles in application 1, we have the 6 × 6 Fisher matrix with elements:

$$I_{F}^{dip} = \begin{bmatrix} I_{F,\alpha_{1}\alpha_{1}}^{dip} & I_{F,\alpha_{1}\mathbf{r}_{\perp,1}}^{dip} & I_{F,\alpha_{1}\alpha_{2}}^{dip} & I_{F,\alpha_{1}\mathbf{r}_{\perp,2}}^{dip} \\ I_{F,\mathbf{r}_{\perp,1}\alpha_{1}}^{dip} & I_{F,\mathbf{r}_{\perp,1}\mathbf{r}_{\perp,1}}^{dip} & I_{F,\mathbf{r}_{\perp,1}\alpha_{2}}^{dip} & I_{F,\mathbf{r}_{\perp,1}\mathbf{r}_{\perp,2}}^{dip} \\ I_{F,\alpha_{2}\alpha_{1}}^{dip} & I_{F,\alpha_{2}\mathbf{r}_{\perp,1}}^{dip} & I_{F,\alpha_{2}\alpha_{2}}^{dip} & I_{F,\alpha_{2}\mathbf{r}_{\perp,2}}^{dip} \\ I_{F,\alpha_{2}\alpha_{1}}^{dip} & I_{F,\alpha_{2}\mathbf{r}_{\perp,1}}^{dip} & I_{F,\alpha_{2}\alpha_{2}}^{dip} & I_{F,\alpha_{2}\mathbf{r}_{\perp,2}}^{dip} \end{bmatrix},$$
(33)

where $I_{F,\mathbf{r}_{\perp,i}\mathbf{r}_{\perp,i}}^{\text{dip}}$, $I_{F,\mathbf{r}_{\perp,i}\alpha_{\perp,i}}^{\text{dip}}$ and $I_{F,\alpha_{\perp,i}\mathbf{r}_{\perp,i}}^{\text{dip}}$ are 2 × 2, 2 × 1 and 1 × 2 sub-matrices, respectively. The diagonal elements of I_F^{dip} are:

$$I_{F,\alpha_{i}\alpha_{i}}^{\text{dip}} = \frac{2}{\hbar\omega} \sum_{\mathbf{r}_{\perp}',j} \frac{\left|\mathcal{F}^{-1}\left(\boldsymbol{\Psi}_{j,i}\right)\left(\mathbf{r}_{\perp}'\right)\right|^{2}}{\alpha_{i}^{2}} + \frac{2}{\hbar\omega} \sum_{\mathbf{r}_{\perp}',j} \Re\left(\frac{\mathcal{F}^{-1}\left(\boldsymbol{\Psi}_{j}\right)^{*}\left(\mathbf{r}_{\perp}'\right)\left[\mathcal{F}^{-1}\left(\boldsymbol{\Psi}_{j,i}\right)\left(\mathbf{r}_{\perp}'\right)\right]^{2}}{\alpha_{i}^{2}\mathcal{F}^{-1}\left(\boldsymbol{\Psi}_{j}\right)\left(\mathbf{r}_{\perp}'\right)}\right), \quad (34)$$

$$\overset{\text{lip}}{=} \frac{2}{2} \sum_{\mathbf{r}_{\perp}'} \sum_{i} \left|\nabla_{\mathbf{r}_{\perp}'}\mathcal{F}^{-1}\left(\boldsymbol{\Psi}_{\perp}\right)\left(\mathbf{r}_{\perp}'\right)\right|^{2}$$

$$I_{F,\mathbf{r}_{\perp,i}\mathbf{r}_{\perp,i}}^{\text{dip}} = \frac{2}{\hbar\omega} \sum_{\mathbf{r}_{\perp}',j} \left| \nabla_{\mathbf{r}_{\perp}'} \mathcal{F}^{-1} \left(\Psi_{j,i} \right) (\mathbf{r}_{\perp}') \right|^{2} + \frac{2}{\hbar\omega} \sum_{\mathbf{r}_{\perp}',j} \Re\left(\frac{\mathcal{F}^{-1} \left(\Psi_{j} \right)^{*} (\mathbf{r}_{\perp}') \left[\nabla_{\mathbf{r}_{\perp}'} \mathcal{F}^{-1} \left(\Psi_{j,i} \right) (\mathbf{r}_{\perp}') \right]^{2}}{\mathcal{F}^{-1} \left(\Psi_{j} \right) (\mathbf{r}_{\perp}')} \right|, \quad (35)$$

which are given by Eq. (S17) and Eq. (S18) in the supplementary document.

It is of interest to first study the diagonal terms in I_F . For instance, suppose that we have exact knowledge about the illumination power, the first dipole's position and the second dipole's strength and position, then $(I_{F,\alpha_1\alpha_1}^{\text{dip}})^{-1}$ is the CRLB of α_1 for any unbiased estimator. When only one dipole exists in the sample, the diagonal terms in I_F^{dip} can be rewritten as:

$$I_{F,\alpha_{1}\alpha_{1}}^{\text{dip}} = \frac{4}{\hbar\omega} \sum_{\mathbf{r}_{\perp}',j} \left| \mathcal{F}^{-1} \left[Q(\mathbf{k}_{\perp} + \mathbf{k}_{\perp,j}) e^{-i\mathbf{k}_{\perp} \cdot \mathbf{r}_{\perp,1}} \right] \right|^{2},$$
(36)

$$I_{F,\mathbf{r}_{\perp,1}\mathbf{r}_{\perp,1}}^{\mathrm{dip}} = \frac{4}{\hbar\omega} \left| C_{1}^{\mathrm{dip}} \right|^{2} \sum_{\mathbf{r}_{\perp}',j} \left[\frac{J_{2} \left(k \mathrm{NA} \left| \mathbf{r}_{\perp}' - \mathbf{r}_{\perp,1} \right| \right)^{2}}{\left| \mathbf{r}_{\perp}' - \mathbf{r}_{\perp,1} \right|^{2}} \right],$$
(37)

where C_1^{dip} is the complex valued constant:

$$C_1^{\rm dip} = \frac{\alpha_1 A_{\rm in} k^4 \rm N A^2}{8i\pi\epsilon_0 k_z}.$$
(38)

In Eq. (37) we used the following relation [54]:

$$\frac{d}{dx}\left(\frac{J_1(x)}{x}\right) = \frac{-J_2(x)}{x},\tag{39}$$

where J_1 and J_2 are the Bessel function of the first kind of order 1 and 2, respectively.

We can see in Eq. (36) that the CRLB of α_i is inversely proportional to the total illumination power A_{in}^2 . Therefore, it is needed to enhance the illumination power to determine the value of α_i for smaller particles. However, when the illumination power is enhanced too much, one may reach a saturation point due to the limited dynamic range of the detector. By taking dark field images of the sample, as shown in Fig. 3, one can avoid this limit. Furthermore, we observe that Table 5

The variance and squared bias of a_1 and x_1 of the dipole of i = 1 for different photon counts PN^{dip}, obtained from Monte Carlo result.

PN ^{dip}	10 ⁴	106	10 ⁸
$\operatorname{Var}\left[\alpha_{i=1}/(\lambda^3)\right]$	3.14×10^{-12}	2.54×10^{-14}	2.62×10^{-16}
Bias $\left[\alpha_{i=1}/(\lambda^3)\right]^2$	3.22×10^{-10}	1.03×10^{-13}	1.26×10^{-17}
$\operatorname{Var}\left(x_{i=1}/\lambda\right)$	4.54×10^{-6}	4.28×10^{-8}	4.23×10^{-10}
Bias $(x_{i=1}/\lambda)^2$	4.32×10^{-7}	4.03×10^{-11}	1.89×10^{-13}

 $I_{F,\mathbf{r}_{\perp,1}\mathbf{r}_{\perp,1}}^{\text{dip}}$ does not only depend on the values of *A* and α_1 , but also on the NA. Therefore, to decrease the CRLB of $\mathbf{r}_{\perp,1}$, one can increase the illumination power or one can enlarge NA, or enhance both. It is interesting that $I_{F,\mathbf{r}_{\perp,1}\mathbf{r}_{\perp,1}}^{\text{dip}}$ is not a function of $\mathbf{k}_{\perp,j}$, which indicate that adjusting the illumination's incident angle can lead to any change of the CRLB of $\mathbf{r}_{\perp,1}$ for the case of a single particle.

When more than one particle is on the planar surface, we have to calculate the Fisher information by Eq. (33). We see from these equations that there is a correlation between the particles. Suppose there are two particles, then the CRLB of one of the particles is a function of the parameters of the other particle, as follows from Eq. (34) and (35) where the second terms consist of the complete field Ψ_{j} instead of only the partial field $\Psi_{j,i}$. A more detailed study of the cross-correlation is presented in the next section.

5.2. The CRLB of the dipole

We study the CRLB of the dipole strength and the position of the dipole along the *x*-axis. We follow the configuration as described in Fig. 3 and Table 1. We first investigate the variance and the squared bias of parameters, α_1 and x_1 , of the dipole i = 1. To find the variance and bias for various noise levels, we define the illumination power by counting the time-averaged number of photons scattered by the dipole i = 1, which is given by:

$$PN^{dip} = \frac{\left\| \mathcal{F}^{-1} \left(\frac{A_{in}k^2}{8i\pi\epsilon_0 k_z} \alpha_i e^{-i\mathbf{r}_{\perp,i}\cdot\mathbf{k}_{\perp}} \right) \right\|_{i=1}^2}{\hbar\omega}.$$
 (40)

The variance and bias are obtained from Monte Carlo simulations. We generated 1000 Fourier ptychographic dark field data-sets for $PN^{dip} = 10^4, 10^6, 10^8$. The parameters are reconstructed from the data-sets by applying the parameter retrieval algorithm described in Section 2.4. The variance and squared bias for $PN^{dip} = 10^4, 10^6, 10^8$, are shown in Table 5.

When $PN^{dip} = 10^4$, we see that the variance of x_1 obtained from the retrieval method is 10 times larger than the squared bias. This variancebias-ratio becomes higher when PN^{dip} is increased. This observation means that the retrieval method of x_1 is asymptotically unbiased when $PN^{dip} > 10^4$. These variances are illustrated in Fig. 8, together with the computed CRLB. It is shown that the variance of the retrieval of x_1 is indeed bounded by the CRLB when $PN^{dip} > 10^4$. The value of the bound is inversely proportional to the value of PN^{dip} .

However, Table 5 also shows that the variance of α_1 obtained from the algorithm is much smaller than the squared bias when PN^{dip} < 10⁶, and indeed the retrieval algorithm of α_1 is not unbiased as long as PN^{dip} < 10⁸ for the current setup. Therefore, the variance of the retrieved α_i may not be bounded by the CRLB when PN^{dip} < 10⁸. On the other hand, we can see in Eq. (14) that the accuracy of the reconstruction of α_i is not only influenced by the Poisson noise, but also by the fluctuation of the illumination power A_{in}^2 . That is, the uncertainty about the exact value of *A* will lead to uncertainty of the retrieval of α_1 . Therefore, it is more difficult to determine α_1 than the position with the current scheme.

5.3. The correlation between two dipoles

As has been noted in Section 5.1, when there are two particles on the surface, varying the parameters of one particle can lead to a change of the CRLB of another particle. To verify this correlation between the particles, we calculated the CRLB of x_1 with various values of α_2 . The value of PN^{dip} is chosen to be 10⁸ because the retrieval algorithm is asymptotically unbiased for this noise level, as has been shown in Section 5.2. The computed CRLB is validated by using Monte Carlo simulations, as illustrated in Fig. 8(b).

It is seen in Fig. 8(b) that one can lower the CRLB of x_1 obtained from the algorithm by enhancing the scattering power of the dipole i = 2. This observation can be understood by studying the property of the Poisson distribution. The signal-to-noise ratio (SNR) of Poisson noise is equal to $\sqrt{n(\mathbf{r}'_1)}$, where $n(\mathbf{r}'_1)$ is the number of photons detected by the pixel at \mathbf{r}'_1 . When the scattering power of particle i = 1 is fixed, $n(\mathbf{r}'_1)$ is increased by enhancing the scattering power of the other particle, and therefore the signal-to-noise ratio of the system is increased. One may argue that this conclusion is inconsistent with the case where incoherent illumination is used. Let us imagine that we apply incoherent illumination to the setup in Fig. 2, then the radiation of each dipole is independent to the other and hence the image recorded by the detector is given by:

$$I_{j}^{\text{incoh}}(\mathbf{r}_{\perp}') = \sum_{i} \left| \mathcal{F}^{-1} \left[\alpha_{i} Q(\mathbf{k}_{\perp} + \mathbf{k}_{\perp,j}) \right] \right|^{2} (\mathbf{r}_{\perp}' - \mathbf{r}_{\perp,i}')$$
$$= \sum_{i} I_{j,i}^{\text{incoh}}(\mathbf{r}_{\perp}').$$
(41)

When there are two dipoles, Eq. (41) shows that the signal of dipole i = 1 is $I_{j,1}^{\text{incoh}}(\mathbf{r}'_{\perp})$ whereas the variance of the signal is determined by $\sum_i I_{j,i}^{\text{incoh}}(\mathbf{r}'_{\perp})$ at the neighborhood of position $\mathbf{r}_{\perp,1}$. Therefore, for the case of incoherent illumination, the SNR of dipole i = 1 should be decreased by enhancing the scattering power of the dipole i = 2 because the variance is proportional to $\sum_i I_{j,i}^{\text{incoh}}(\mathbf{r}'_{\perp})$ for Poisson noise. However, we emphasize that Eq. (41) is not the case of application 1. By comparing Eq. (22) to Eq. (41), we see that the measurement in application 1 contains the interference pattern of the point spread function of the dipoles. Hence, the conclusion for incoherent illumination is not applicable in application 1 and the SNR should be determined by the computed CRLB and the Monte Carlo simulations. Note that second order scattering is neglected in the current model, i.e. we ignore the scattered wavefield from the first particle which is excited by the second one because the particles are sparsely distributed on the sample.

6. The CRLB analysis of application 2

6.1. Fisher Information matrix for single rectangular object

For application 2, the parameter vector we want to retrieve is:

$$\boldsymbol{\Theta} = \begin{bmatrix} \theta_1, \theta_2, \ldots \end{bmatrix}^T = \begin{bmatrix} A_1, \phi_1, a_1, b_1, \mathbf{r}_{\perp,1} \end{bmatrix}^T,$$
(42)

To find the Fisher information matrix, we start from the expectation of the second order perturbation of \mathcal{L}_P :

$$E\left(\delta^{2}\mathcal{L}_{P}\right)\left(\Theta\right)\left(\delta\Theta,\delta\Theta\right)$$

$$=\frac{2}{\hbar\omega}\sum_{\mathbf{k}_{\perp}',j}\Re\left[\mathcal{F}\left[P_{j}\delta O(\Theta)(\delta\Theta)\right]\mathcal{F}\left[P_{j}\delta O(\Theta)(\delta\tilde{\Theta})\right]^{*}\right]$$

$$+\frac{2}{\hbar\omega}\sum_{\mathbf{k}_{\perp}',j}\Re\left[\frac{\mathcal{F}\left(\Psi_{j}\right)}{\mathcal{F}\left(\Psi_{j}\right)^{*}}\mathcal{F}\left[P_{j}\delta O(\Theta)(\delta\Theta)\right]^{*}\mathcal{F}\left[P_{j}\delta O(\Theta)(\delta\tilde{\Theta})\right]^{*}\right].$$
(43)

which is derived from Eq. (S11) in Supplement. The function *O* is defined in Eq. (26). δO is the derivative of *O* w.r.t. Θ . $\delta \Theta$ and $\delta \tilde{\Theta}$ are small perturbations of the parameters of the rectangle. The explicit expression of δO , $\delta \Theta$ and $\delta \tilde{\Theta}$ are given in Section 1.D of the Supplementary document.



Fig. 8. (a) The computed CRLB and variance of the position of the first dipole, i.e. x_1 , for various PN^{dip}. (b) The computed CRLB and variance of x_1 for various values of the polarizability of the second dipole, i.e. a_2 , for the case of PN^{dip} = 10⁸. The blue line of both plots are the computed CRLB and the red crosses show the variance obtained from the Monte Carlo experiment.

By using Eq. (43), Eq. (S2) and Eq. (S23) in the Supplement, we obtain the diagonal elements of the Fisher matrix:

$$I_{F,A_{1}A_{1}}^{\text{rect}} = \frac{2}{\hbar\omega} \sum_{\mathbf{r},j} \left| P_{j} \Pi_{a_{1},b_{1},\mathbf{r}_{1}} \right|^{2} + \frac{2}{\hbar\omega} \sum_{\mathbf{r},j} \Re \left[\mathcal{F}^{-1} \left(\frac{\mathcal{F} \left(\Psi_{j} \right)}{\mathcal{F} \left(\Psi_{j} \right)^{*}} \right) e^{-2i\phi_{1}} \left[\left(P_{j} \Pi_{a_{1},b_{1},\mathbf{r}_{1}} \right)^{*} \right]^{2} \right].$$
(44)

$$I_{F,\phi_1\phi_1}^{\text{rect}} = A_1^2 I_{F,A_1A_1}.$$
 (45)

$$\begin{split} I_{F,a_{1}a_{1}}^{\text{rect}} &= \frac{1}{2\hbar\omega} \sum_{y,j} \left| C_{1} \Pi_{b_{1},y_{1}} \right|^{2} \left| \left| P_{j} \right|^{2} (x_{1} + \frac{a_{1}}{2}, y) + \left| P_{j} \right|^{2} (x_{1} - \frac{a_{1}}{2}, y) \right| \\ &+ \frac{1}{2\hbar\omega} \sum_{y,j} \Re \left[\left(C_{1}^{*} \Pi_{b_{1},y_{1}} \right)^{2} \mathcal{F}^{-1} \left(\frac{\mathcal{F} \left(\Psi_{j} \right)}{\mathcal{F} \left(\Psi_{j} \right)^{*}} \right) \right. \\ &\times (2x_{1} + a_{1}, y) \left(P_{j}^{*} \right)^{2} (x_{1} + \frac{a_{1}}{2}, y) \right] \\ &+ \frac{1}{2\hbar\omega} \sum_{y,j} \Re \left[\left(C_{1}^{*} \Pi_{b_{1},y_{1}} \right)^{2} \mathcal{F}^{-1} \left(\frac{\mathcal{F} \left(\Psi_{j} \right)}{\mathcal{F} \left(\Psi_{j} \right)^{*}} \right) \right. \\ &\times (2x_{1} - a_{1}, y) \left(P_{j}^{*} \right)^{2} (x_{1} - \frac{a_{1}}{2}, y) \right] \\ &+ \frac{1}{\hbar\omega} \sum_{y,j} \Re \left[\left(C_{1}^{*} \Pi_{b_{1},y_{1}} \right)^{2} \mathcal{F}^{-1} \left(\frac{\mathcal{F} \left(\Psi_{j} \right)}{\mathcal{F} \left(\Psi_{j} \right)^{*}} \right) \right. \\ &\times (2x_{1}, y) P_{j}^{*} (x_{1} + \frac{a_{1}}{2}, y) P_{j}^{*} (x_{1} - \frac{a_{1}}{2}, y) \right]. \end{split}$$
(46)

 $I_{F,b_1b_1}^{\text{rect}}$ can be obtained by taking the above equation and interchanging *x* with *y* and *a*₁ with *b*₁.

$$\begin{aligned} f_{F,x_{1}x_{1}}^{\text{rect}} &= \frac{2}{\hbar\omega} \sum_{y,j} \left| C_{1} \Pi_{b_{1},y_{1}} \right|^{2} \left[\left| P_{j} \right|^{2} (x_{1} + \frac{a_{1}}{2}, y) + \left| P_{j} \right|^{2} (x_{1} - \frac{a_{1}}{2}, y) \right] \\ &+ \frac{2}{\hbar\omega} \sum_{y,j} \Re \left[\left(C_{1}^{*} \Pi_{b_{1},y_{1}} \right)^{2} \mathcal{F}^{-1} \left(\frac{\mathcal{F} \left(\Psi_{j} \right)}{\mathcal{F} \left(\Psi_{j} \right)^{*}} \right) \right] \\ &\times (2x_{1} + a_{1}, y) \left(P_{j}^{*} \right)^{2} (x_{1} + \frac{a_{1}}{2}, y) \right] \\ &+ \frac{2}{\hbar\omega} \sum_{y,j} \Re \left[\left(C_{1}^{*} \Pi_{b_{1},y_{1}} \right)^{2} \mathcal{F}^{-1} \left(\frac{\mathcal{F} \left(\Psi_{j} \right)}{\mathcal{F} \left(\Psi_{j} \right)^{*}} \right) \right] \\ &\times (2x_{1} - a_{1}, y) \left(P_{j}^{*} \right)^{2} (x_{1} - \frac{a_{1}}{2}, y) \right] \\ &- \frac{4}{\hbar\omega} \sum_{y,j} \Re \left[\left(C_{1}^{*} \Pi_{b_{1},y_{1}} \right)^{2} \mathcal{F}^{-1} \left(\frac{\mathcal{F} \left(\Psi_{j} \right)}{\mathcal{F} \left(\Psi_{j} \right)^{*}} \right) \right] \\ &\times (2x_{1}, y) P_{j}^{*} (x_{1} + \frac{a_{1}}{2}, y) P_{j}^{*} (x_{1} - \frac{a_{1}}{2}, y) \right]. \end{aligned}$$

 $I_{F,y_1y_1}^{\text{rect}}$ can be obtained by taking the above equation and interchanging *x* with *y* and a_1 with b_1 .

We again focus on the diagonal elements of the Fisher matrix. Referring to the first term on the right-hand side of Eqs. (44) and (45), we can immediately see that the CRLB of A_1 and ϕ_1 is partially determined by the illumination power. Similarly, in Eqs. (46) and (47) we see that the CRLB of a_1 and x_1 is partially determined by the illumination power at $x_1 \pm \frac{a_1}{2}$, which is the edge of the rectangle. We can also notice in Eq. (45) that the CRLB of ϕ_1 is inversely proportional to A_1^2 . This observation means that one can retrieve ϕ_1 more accurately by increasing the transmission of the rectangle, assuming that the estimator is unbiased.

It is interesting that $I_{F,a_1a_1}^{\text{rect}}$ and $I_{F,x_1x_1}^{\text{rect}}$ are functions of Π_{b_1,y_1} . This fact means that enlarging the width of the rectangle in the *y*-direction will decrease the CRLB of a_1 and x_1 , which are parameters along the *x*-axis. This correlation between b_1 and the CRLB of a_1 and x_1 is demonstrated in the next subsection. The computed CRLB is validated by Monte Carlo simulations.

6.2. The CRLB of the width and the position of the rectangle

Now we consider the configuration of Section 3. As described in Section 5.2, we need to provide a measure of the noise level in terms of photon counting. For application 2, we define the illumination power by means of the total photon number counting over the cross section of the probe:

$$PN^{rect} = \sum_{\mathbf{r}_{\perp}}^{N_x^{det}, N_y^{det}} \frac{\|P(\mathbf{r}_{\perp})\|^2}{\hbar\omega},$$
(48)

where the probe $P(\mathbf{r}_{\perp})$ is shown in Fig. 6(b). We note that the summation in Eq. (48) is from 1 to N_x^{det} in the *x* direction and is from 1 to N_y^{det} in the *y* direction, which is due to the grid size of the probe is the same as the grid size of the detector array.

Here we study the influence of the width of the rectangle in the *y*-direction on the variance of retrieved width and position along the *x*-axis. The computed CRLB of a_1 and x_1 are shown in Fig. 9, for various values of b_1 . The value of PN^{rect} is chosen to be 10⁸. To validate the computation of the CRLB, the result of Monte Carlo Monte simulations is shown in Fig. 9 also. To obtain the variance, 1000 ptychographic data-sets are created in the Monte Carlo analysis. The data-sets are post-processed by using the parameter retrieval algorithm given in Section 3.2. The exact value of the variance and the squared bias of the parameters for the case of $b_1/\lambda = 1, 5, 15$, are listed in Table 6.

We see in Fig. 9(a) and Fig. 9(b) that when $PN^{rect} = 10^8$ the CRLB of a_1/λ and x_1/λ are in the order of 10^{-6} , which indicates that the resolution of the current parameter retrieval scheme is not limited by the grid discretization in real space. The Monte Carlo result confirm this conclusion. Moreover, the squared bias of a_1/λ and x_1/λ is around 10^3 times smaller that the variance, which means the that algorithm



Fig. 9. The CRLB and variance of a_1 , x_1 , b_1 and y_1 of the rectangle, for various of b_1 . The PN^{rect} of this figure is 10^8 . The blue line is the computed CRLB and the red crosses show the variance obtained from the Monte Carlo simulations.

Table 6 The variance and squared bias of a_1 and x_1 of the rectangle, obtained from the Monte Carlo simulation for PN^{rect} = 10^8 .

Guilo billiulution io								
b_1/λ	1	5	15					
$\operatorname{Var}\left(a_{1}/\lambda\right)$	3.576×10^{-7}	1.455×10^{-7}	9.017×10^{-8}					
Bias $(a_1/\lambda)^2$	7.825×10^{-10}	4.254×10^{-10}	8.386×10^{-11}					
$\operatorname{Var}\left(x_{1}/\lambda\right)$	9.057×10^{-8}	2.527×10^{-8}	1.824×10^{-8}					
Bias $(x_1/\lambda)^2$	6.423×10^{-12}	6.879×10^{-11}	4.947×10^{-11}					

is asymptotically unbiased when $PN^{rect} = 10^8$, and hence the variance obtained by the algorithm should be bounded by the CRLB. The CRLB of both a_1/λ and x_1/λ decrease when the value of b_1 is increased. This result agrees with Eqs. (46) and (47). The CRLB of a_1/λ and x_1/λ in Fig. 9 decreases rapidly when $b_1/\lambda < 5$. The reason is that the sensitivity of the retrieval of the parameters is determined by the number of photons which encodes the information about the parameters. That is, there are more photons which contain information about a_1 and x_1 when b_1 is larger. On the other hand, we can see that the CRLB of b_1/λ and y_1/λ do not vary much when the value of b_1/λ is sufficiently small. When $b_1/\lambda > 40$, the CRLB of b_1/λ and y_1/λ start to increase as the value of b_1/λ is enlarged. This is because the boundary of the rectangle parallel to the y-axis falls outside of the illuminated area, which is an undesirable situation since b_1/λ and y_1/λ need also to be retrieved. To be explicit, the size of the illuminated area is determined by the non-zero area of $\sum_{j} |P(\mathbf{r}_{\perp} - \mathbf{R}_{\perp,j})|^2$. In the simulation, the size of the illuminated area in the *y* direction is roughly 60λ . Meanwhile, the beam profile of the illumination, i.e. $|P(\mathbf{r}_{\perp})|^2$, is simulated by the Gaussian function as shown in Fig. 6(b). The full-width at halfmaximum (FWHM) of the probe is around 15λ . These characteristic parameters of the probe agree with Fig. 9(c) and Fig. 9(d) that the CRLB of b_1/λ and y_1/λ start to increase as $b_1/\lambda > 40$. Overall, the computed CRLB as shown in Fig. 9 indicates that, the optimal chosen range of values of b_1/λ is (5,40) for the current configuration.

7. Conclusion

In summary, a parameter retrieval method is demonstrated in this paper. The idea of the method is to incorporate available *a priori* information about the object in the general ptychography framework. Two applications of the method are studied. In application 1 we explore how the parameters of small particles can be retrieved from Fourier ptychographic dark field measurements. The simulation result shows that, when the diameters of the particles are sufficiently small, e.g. $\sim 0.1\lambda$, so that the scattered wavefields can be modeled as radiation of dipoles, the parameters of the particles can be uniquely determined from dark field measurement only. In application 2 the retrieval of the parameters of a rectangular object embedded in constant surrounding was studied.

The influence of Poisson noise on the parameter retrieval method is discussed in the second part of the paper. The CRLB of the parameters are theoretically derived and numerically computed from the Fisher information matrix for both applications. Monte Carlo analysis is used to validate the computed CRLB. The CRLB, variance and bias of the retrieved parameters in application 1 were determined for various photon counts. It was found that the uncertainty of the parameter retrieval is inversely proportional to the photon counts, and potentially is not limited by the sizes of individual cells of the discretized meshgrid in object space. The correlation between at least two particles is evaluated from the calculation of the CRLB. We proved that the CRLB of the position of one particle is influenced by the scattering power of the other particle. This conclusion is confirmed by the Monte Carlo result. The correlation of parameters in application 2 is also inferred from the computed CRLB. The influence of the width of the rectangle in the y-direction on the CRLB of the parameters along the x-axis is investigated by analyzing the CRLB and the Monte Carlo result. For the same number of photons in the illuminating probe, the uncertainty of the parameters along the *x*-axis can be reduced by enlarging the width in the y-direction.

Following this study, there are many related subjects which deserve further research. For Application 1, we only test the proposed method with PSL beads. In practice, the PSL beads are used as the calibration standard with which we can determine the lower detection limit of the 'RapidNano' particle scanner. Generally speaking, the scanner should be able to detect particles of any material. Therefore, to apply the proposed method to other particles is an interesting subject for the next step. Note that the refraction index of the particle is possibly unknown or/and is not the same for different particle on the substrate. For this case, the model used in the paper is unable to retrieve the volume of the particle. To incorporate the proposed method to this case, a more complex model is needed and the subject is out of the scope of the paper. Another issue in Application 1 relates to the fact that we have not considered the secondary scattering effect between adjacent dipoles. This is because that in practice the particles are sparsely distributed on the substrate for the most of time. When two dipoles are very close to each other, the present scattering model cannot accurately describe the scattered field, and hence multiple scattering has to be considered. The multiple scattering is one of the major topics which deserve further research.

For Application 2, one of the subjects for the next step is to investigate the influence of unknown background transmission. When the background transmission is homogeneous but unknown, it is clear that the background transmission can be retrieved from the reconstructed object that obtained from the MLE ptychographic algorithm. Whereas when there is small unknown perturbation in the background transmission, the effect of the perturbation may also be studied via CRLB analysis or Monte Carlo simulation. This analysis could be conducted when we have more knowledge about how/why can we model the perturbation of background. As far as we know, this subject is still an open question and deserves further research.

See Supplement 1 for supporting content.

Funding

H2020 Marie Skłodowska-Curie Actions (675745).

Acknowledgment

X. Wei thanks O. el Gawhary for fruitful comments.

Appendix A. Supplementary data

Supplementary material related to this article can be found online at https://doi.org/10.1016/j.ultramic.2021.113335.

References

- W. Hoppe, Beugung im inhomogenen Primärstrahlwellenfeld. I. Prinzip einer Phasenmessung von Elektronenbeungungsinterferenzen, Acta Crystallogr. Sect. A 25 (4) (1969) 495–501.
- [2] J.M. Rodenburg, R.H.T. Bates, The theory of super-resolution electron microscopy via Wigner-distribution deconvolution, Phil. Trans. R. Soc. A 339 (1655) (1992) 521–553.
- [3] H.N. Chapman, Phase-retrieval X-ray microscopy by Wigner-distribution deconvolution, Ultramicroscopy 66 (3–4) (1996) 153–172.
- [4] H.M.L. Faulkner, J.M. Rodenburg, Movable aperture lensless transmission microscopy: A novel phase retrieval algorithm, Phys. Rev. Lett. 93 (2) (2004) 023903.
- [5] J.M. Rodenburg, H.M.L. Faulkner, A phase retrieval algorithm for shifting illumination, Appl. Phys. Lett. 85 (20) (2004) 4795–4797.
- [6] M. Guizar-Sicairos, J.R. Fienup, Phase retrieval with transverse translation diversity: a nonlinear optimization approach, Opt. Express 16 (10) (2008) 7264–7278.
- [7] J.C. da Silva, A. Menzel, Elementary signals in ptychography, Opt. Express 23 (26) (2015) 33812–33821.
- [8] M.D. Seaberg, B. Zhang, D.F. Gardner, E.R. Shanblatt, M.M. Murnane, H.C. Kapteyn, D.E. Adams, Tabletop nanometer extreme ultraviolet imaging in an extended reflection mode using coherent Fresnel ptychography, Optica 1 (1) (2014) 39–44.
- [9] M. Odstrcil, J. Bussmann, D. Rudolf, R. Bresenitz, J. Miao, W.S. Brocklesby, L. Juschkin, Ptychographic imaging with a compact gas-discharge plasma extreme ultraviolet light source, Opt. Lett. 40 (23) (2015) 5574–5577.
- [10] J.M. Rodenburg, A.C. Hurst, A.G. Cullis, B.R. Dobson, F. Pfeiffer, O. Bunk, C. David, K. Jefimovs, I. Johnson, Hard-X-Ray lensless imaging of extended objects, Phys. Rev. Lett. 98 (3) (2007) 034801.
- [11] P. Thibault, M. Dierolf, A. Menzel, O. Bunk, C. David, F. Pfeiffer, High-resolution scanning X-ray diffraction microscopy, Science 321 (5887) (2008) 379–382.
- [12] H.N. Chapman, K.A. Nugent, Coherent lensless X-ray imaging, Nat. Photonics 4 (12) (2010) 833–839.
- [13] F. Pfeiffer, X-ray ptychography, Nat. Photonics 12 (1) (2017) 9-17.
- [14] P. Thibault, M. Dierolf, O. Bunk, A. Menzel, F. Pfeiffer, Probe retrieval in ptychographic coherent diffractive imaging, Ultramicroscopy 109 (4) (2009) 338–343.
- [15] A.M. Maiden, J.M. Rodenburg, An improved ptychographical phase retrieval algorithm for diffractive imaging, Ultramicroscopy 109 (10) (2009) 1256–1262.
- [16] M. Holler, M. Guizar-Sicairos, E.H.R. Tsai, R. Dinapoli, E. Müller, O. Bunk, J. Raabe, G. Aeppli, High-resolution non-destructive three-dimensional imaging of integrated circuits, Nature 543 (7645) (2017) 402–406.
- [17] D.F. Gardner, M. Tanksalvala, E.R. Shanblatt, X. Zhang, B.R. Galloway, C.L. Porter, R. Karl Jr., C. Bevis, D.E. Adams, H.C. Kapteyn, M.M. Murnane, G.F. Mancini, Subwavelength coherent imaging of periodic samples using a 13.5 nm tabletop high-harmonic light source, Nat. Photonics 11 (4) (2017) 259–263.
- [18] Y. Jiang, Z. Chen, Y. Han, P. Deb, H. Gao, S. Xie, P. Purohit, M.W. Tate, J. Park, S.M. Gruner, V. Elser, D.A. Muller, Electron ptychography of 2D materials to deep sub-ångström resolution, Nature 559 (7714) (2018) 343–349.
- [19] A. Maiden, D. Johnson, P. Li, Further improvements to the ptychographical iterative engine, Optica 4 (7) (2017) 736–745.
- [20] G. Zheng, R. Horstmeyer, C. Yang, Wide-field, high-resolution Fourier ptychographic microscopy, Nat. Photonics 7 (9) (2013) 739–745.
- [21] X. Ou, G. Zheng, C. Yang, Embedded pupil function recovery for Fourier ptychographic microscopy, Opt. Express 22 (5) (2014) 4960.
- [22] R. Horstmeyer, C. Yang, A phase space model of Fourier ptychographic microscopy, Opt. Express 22 (1) (2014) 338.
- [23] L.-H. Yeh, J. Dong, J. Zhong, L. Tian, M. Chen, G. Tang, M. Soltanolkotabi, L. Waller, Experimental robustness of Fourier ptychography phase retrieval algorithms, Opt. Express 23 (26) (2015) 33214–33240.
- [24] R. Horstmeyer, R.Y. Chen, X. Ou, B. Ames, J.A. Tropp, C. Yang, Solving ptychography with a convex relaxation, New J. Phys. 17 (5) (2015) 053044.

- [25] Y. Zhang, W. Jiang, Q. Dai, Nonlinear optimization approach for Fourier ptychographic microscopy, Opt. Express 23 (26) (2015) 33822.
- [26] V. Elser, Phase retrieval by iterated projections, J. Opt. Soc. Amer. A 20 (1) (2003) 40–55.
- [27] P. Godard, M. Allain, V. Chamard, J. Rodenburg, Noise models for low counting rate coherent diffraction imaging, Opt. Express 20 (23) (2012) 25914–25934.
- [28] P. Thibault, M. Guizar-Sicairos, Maximum-likelihood refinement for coherent diffractive imaging, New J. Phys. 14 (6) (2012) 063004.
- [29] M. Odstrčil, A. Menzel, M. Guizar-Sicairos, Iterative least-squares solver for generalized maximum-likelihood ptychography, Opt. Express 26 (3) (2018) 3108–3123.
- [30] A.M. Maiden, M.J. Humphry, F. Zhang, J.M. Rodenburg, Superresolution imaging via ptychography, J. Opt. Soc. Amer. A 28 (4) (2011) 604.
- [31] A. Szameit, Y. Shechtman, E. Osherovich, E. Bullkich, P. Sidorenko, H. Dana, S. Steiner, E.B. Kley, S. Gazit, T. Cohen-Hyams, S. Shoham, M. Zibulevsky, I. Yavneh, Y.C. Eldar, O. Cohen, M. Segev, Sparsity-based single-shot subwavelength coherent diffractive imaging, Nature Mater. 11 (5) (2012) 455–459.
- [32] P. Sidorenko, O. Kfir, Y. Shechtman, A. Fleischer, Y.C. Eldar, M. Segev, O. Cohen, Sparsity-based super-resolved coherent diffraction imaging of one-dimensional objects, Nature Commun. 6 (1) (2015).
- [33] D. Slepian, H.O. Pollak, Prolate spheroidal wave functions, fourier analysis and uncertainty - I, Bell Syst. Tech. J. 40 (1) (1961) 43–63.
- [34] A. Papoulis, A new algorithm in spectral analysis and band-limited extrapolation, IEEE Trans. Circuits Syst. 22 (9) (1975) 735–742.
- [35] P. Delsarte, A.J.E.M. Janssen, L.B. Vries, Discrete prolate spheroidal wave functions and interpolation, SIAM J. Appl. Math. 45 (4) (1985) 641–650.
- [36] P. van der Walle, S. Hannemann, D. van Eijk, W. Mulckhuyse, J.C.J. van der Donck, Implementation of background scattering variance reduction on the rapid nano particle scanner, in: J.P. Cain, M.I. Sanchez (Eds.), Metrology, Inspection, and Process Control for Microlithography XXVIII, SPIE, 2014, http://dx.doi.org/ 10.1117/12.2048684.
- [37] P. van der Walle, E. Kramer, J.C.J. van der Donck, W. Mulckhuyse, L. Nijsten, F.A.B. Arango, A. de Jong, E. van Zeijl, H.E.T. Spruit, J.H. van den Berg, G. Nanda, A.K. van Langen-Suurling, P.F.A. Alkemade, S.F. Pereira, D.J. Maas, Deep sub-wavelength metrology for advanced defect classification, in: P. Lehmann, W. Osten, A.A. Gonçalves (Eds.), Optical Measurement Systems for Industrial Inspection X, SPIE, 2017, http://dx.doi.org/10.1117/12.2272414.
- [38] P.J.S. Ferreira, The stability of a procedure for the recovery of lost samples in band-limited signals, Signal Process. 40 (2–3) (1994) 195–205.
- [39] A.J. den Boef, Optical wafer metrology sensors for process-robust CD and overlay control in semiconductor device manufacturing, Surf. Topogr.: Metrol. Prop. 4 (2) (2016) 023001.
- [40] P. Ansuinelli, W.M.J. Coene, H.P. Urbach, Automatic feature selection in EUV scatterometry, Appl. Opt. 58 (22) (2019) 5916.
- [41] D.J. Griffiths, Introduction to Electrodynamics, third ed., Prentice Hall, 1999.
- [42] L. Novotny, B. Hecht, Principles of Nano-Optics, Cambridge University Press, 2012.
- [43] J. Goodman, Introduction to Fourier Optics, in: McGraw-Hill Physical and Quantum Electronics Series, W. H. Freeman, 2005.
- [44] P. van der Walle, P. Kumar, D. Ityaksov, R. Versluis, D.J. Maas, O. Kievit, J. Janssen, J.C.J. van der Donck, Increased particle inspection sensitivity by reduction of background scatter variance, in: A. Starikov, J.P. Cain (Eds.), Metrology, Inspection, and Process Control for Microlithography XXVII, SPIE, 2013, http://dx.doi.org/10.1117/12.2011092.
- [45] W. Murray, M.H. Wright, P.E. Gill, Practical Optimization, Emerald Publishing Limited, 1982.
- [46] R. Fletcher, Practical Methods of Optimization, second ed., Wiley, 1988.
- [47] X. Wei, H.P. Urbach, W.M.J. Coene, Cramér-Rao lower bound and maximumlikelihood estimation in ptychography with Poisson noise, Phys. Rev. A 102 (4) (2020).
- [48] O. Bunk, M. Dierolf, S. Kynde, I. Johnson, O. Marti, F. Pfeiffer, Influence of the overlap parameter on the convergence of the ptychographical iterative engine, Ultramicroscopy 108 (5) (2008) 481–487.
- [49] X. Wei, P. Urbach, Ptychography with multiple wavelength illumination, Opt. Express 27 (25) (2019) 36767–36789.
- [50] S. Marchesini, A. Schirotzek, C. Yang, H. tieng Wu, F. Maia, Augmented projections for ptychographic imaging, Inverse Problems 29 (11) (2013) 115009.
- [51] J. Zhong, L. Tian, P. Varma, L. Waller, Nonlinear optimization algorithm for partially coherent phase retrieval and source recovery, IEEE Trans. Comput. Imag. 2 (3) (2016) 310–322.
- [52] S.M. Kay, Fundamentals of Statistical Signal Processing, Volume 1: Estimation Theory, Pearson, 2009.
- [53] D. Bouchet, R. Carminati, A.P. Mosk, Influence of the local scattering environment on the localization precision of single particles, Phys. Rev. Lett. 124 (13) (2020).
- [54] G.B. Arfken, H.J. Weber, Mathematical Methods for Physicists, Elsevier LTD, Oxford, 2012.